



Positive Control for SARS-CoV-2 Serology Assay (SCSA) Catalog Number: AU2022

AU2022 Positive Control for SARS-CoV-2 Serology Assay (SCSA)

Volume: 100 μ l (10 μ L per test)

Storage Conditions: Upon arrival, aliquot and store at -20°C

Supplied In: Antibodies are delivered in 1% BSA/PBS and this may be used as a diluent

Product Description:

Mixture of recombinant humanized IgG and IgM monoclonal antibodies directed against the NP protein of SARS-CoV-2 virus. This serves as an excellent and harmless positive control for Anti-SARS-CoV-2 immunoassays.

Amount recommended for an individual test is 10 μ l.

Example:

The figure 1 below a SARS-CoV-2 antibody LFA assay, the negative control (an irrelevant humanized antibody) is shown in the first panel and the application of AU2022 is shown for Attogene's SARS-CoV-2 Serologic Assay (SCSA) LFA assays.



Figure 1

Related products:

AU2019-XX SARS-CoV-2 Serologic Assay