



Okadaic Acid Lateral Flow Kit (Rapid Lab)

Catalog Number: AU2072-01

For Research Use Only. Not for use in Diagnostic Procedures.

1. Intended Use

For the screening of Okadaic Acid in fresh and seawater samples at 0.5 ppb. Samples requiring regulatory action should be confirmed by ELISA, HPLC, or other conventional methods.

2. Introduction

Attogene's Okadaic Lateral Flow Kit can be used to detect Okadaic Acid in liquid samples. This kit is compatible with fresh and seawater samples.

Format: Rapid-Water - Run Time: 20 Minutes

Okadaic Acid (OA) is a one of the diarrhetic shellfish poisons (DSP) produced by dinoflagellate genera *Dinophysis* and *Prorocentrum*. There are several chemically different toxins associated with DSP. It inhibits key protein phosphatases (PPI and PP2A), causing intestinal cells to leak fluids and leading to severe gastrointestinal distress. Filter-feeding shellfish—such as mussels, clams, and oysters—concentrate the toxin by ingesting the algae, posing a risk to human seafood consumers.

They are lipophilic and polyether compounds and can be divided into three main groups:

1. Acidic toxins
 2. Neutral toxins
 3. Other toxins
- 2 Contamination of shellfish with OA has been associated with harmful algae blooms throughout the world.

3. Kit Contents

Component Name	Volume	Storage
Okadaic Acid Cassette	10 each	RT
Sample Dilution Buffer	5 each	RT
Negative Control	5 each	RT
Manual	1 each	RT

4. Storage and Stability

- The kit should be stored at 2°C - 30°C until ready to use.
- The test must remain in the sealed pouch until use.

5. Required Materials Not Supplied

- Timer - For timing use
- Centrifuge - For preparation of clear specimens
- Pipettor and pipette tips — to transfer samples and controls

6. Precautions

- The Okadaic Acid Lateral Flow Kit provides only preliminary qualitative test results. Use another, more quantitative, analytical method such as ELISA or instrumental analysis to obtain a confirmed quantitative analytical result.
- Prior to use, ensure that the product has not expired by verifying that the date of use is prior to the expiration date on the label.
- The test cassettes are individually packaged in a foil pouch with a desiccant & disposable pipette.
- Avoid cross-contamination of samples by using a new bottle for each sample.
- Use only Okadaic Acid Lateral Flow Kit reagents from one kit lot, as they have been adjusted in combination.
- It is good laboratory practice to use positive and negative controls to ensure proper test performance. Samples which do not contain Okadaic Acid (negative controls) as well as samples containing known quantities of Okadaic Acid (positive controls) should be analyzed with each lot of test strips to provide a reference for line intensity to be expected.

7. Water Collection, Storage and Cell Lysis

- Using gloves, collect water samples into a Water Sample Bottle and store refrigerated for up to 5 days (label the water collection bottle). If samples must be held for greater than 5 days, samples should be stored frozen.
- If cell lysis is required, we recommend performing either 2-3 freeze thaw cycles or sonicating the sample for 3 x 10-60 seconds prior to filtering or centrifuging.
- Allow the test cassettes, running buffer, and samples to reach room temperature before testing.

8. Procedure

- Open two sealed cassettes. It is recommended to perform steps A and B for each sample evaluation.

A. Sample:

NOTE: Each sample dilution buffer tube contains 100ul of a buffering solution necessary for proper running of the strip. Perform a quick centrifugation of the sample dilution buffer tube to ensure the buffer solution is at the bottom of the tube and not stuck on the sides or top.

- Using a pipet transfer 100ul of sample into one of the sample dilution buffer tubes.
- Mix the sample with the sample buffering solution by pipetting up and down 3 times.
- Using a pipet add the buffered sample directly into the sample port of the cassette.

B. Negative Control:

- Pipet 200ul of the solution in the Negative Control tube directly into the sample port of the cassette.

C. Immediately following steps, A and B

- Set a timer for 15-20 minutes.

Read results using your eye or with a lateral flow reader.9. Interpretation of Results

For samples prepared as described above, screening concentrations are determined by comparison of the intensity of the test line to the intensity of the control line on parallel test strips. Although control line intensity may vary, a visible control line must be present for results to be considered valid. Test strips with a test line which is darker than or of equal intensity to the test line of the control indicates a result which is below the limit of detection of the test. Test strips with a test line which is lighter than the control strip indicates a result which is above 0.5 ppb. Results should be determined within 5 minutes after completion of the strip test procedure (20-25 minutes). Determination made using strips which have dried for more or less than the required time may be inaccurate, as line intensities may vary with drying time.

To obtain semi-quantitative results in the range of 0 — 10 ppb, solutions of known Okadaic Acid concentration (control solutions) must be tested concurrently with samples. Sample test line intensities can then be compared with control solution test line intensities, yielding approximate sample concentrations. Do not use strips run previously to determine semi-quantitative sample concentrations, as test line intensities may vary once strips are completely dry.

10. Additional Analysis

If necessary, positive samples can be confirmed by ELISA, HPLC, or other conventional methods. A lateral flow reader may also be employed to generate numerical readings from the visual result. Contact us if you have any questions.

Who we are

Attogene is a biotechnology company located in Austin, Texas. Our focus is to enhance health and wellness by offering and developing customer focused Life Science Products domestically and internationally.

Our mission is to:

- Enhance detection technologies
- Enable rapid responses
- Enable impactful research discoveries

Contact Us

3913 Todd Lane, Suite 310
Austin, TX 78744

Phone: 512- 333-1330

Email: sales@attogene.com

Web: www.attogene.com

AU2024_V5